


Certificate



Certificate no.

CU 72041278 01

License Holder:

Sierra Monitor Corporation
1991 Tarob Court
Milpitas CA 95035
USA

Manufacturing Plant:

Sierra Monitor Corporation
1991 Tarob Court
Milpitas CA 95035
USA

Test report no.: USA-PSM 30481229 001

Client Reference: Michael C. Farr

Tested to: UL 61010-1:2004
CAN/CSA-C22.2 61010-1:2004

Certified Product: Gas Sensor Monitor

License Fee - Units

Model Designation: 20X Series; 200X Series
(X=0-9, not safety-relevant)

7


Rated Voltage: DC 9-24V
Rated Current: 250mA
Protection Class: III

Special Remarks: The units are "Classified" for electrical safety only and only for ordinary, not hazardous locations. They have not been evaluated for performance or reliability.

7

Appendix: 1

Licensed Test mark:



Signatures

Stephan Schmitt
President

Dipl.-Ing. M. Raap
QA Certification Officer

Date of Issue
(day/mo/yr)
10/11/2004