
<CALIBRATION REPORT>

ID:A USER # =0 12/18 10:03

GAS TYPE : COMB
CALIB CONC= 60
CAL FREQ= 90 DAYS
GLOBAL CALIB: OFF

>MODULE # = 2
% CHANGE = +01 %FS
PEAK CONC. = 61
1ST VOLTAGE = +0.073
2ND VOLTAGE = +0.810

>MODULE # = 3
% CHANGE = -15 %FS
PEAK CONC. = 45
1ST VOLTAGE = +1.649
2ND VOLTAGE = +1.395

>MODULE # = 4
LOW SENSIVITY
% CHANGE = +01 %FS
PEAK CONC. = 61
1ST VOLTAGE = +0.073
2ND VOLTAGE = +0.816