

-----  
<HISTORY REPORT>

ID:A USER # =1 12/17 15:08

HISTORY RESET AT: 02/01 10:59  
POWER DOWN AT:12/17 13:34  
POWER UP AT:12/17 13:44

MODULE 1 PPM H2S

LOW ALARM 12/13 05:07  
HIGH ALARM 12/17 15:07  
LOWEST CONC = 00 12/16 10:59  
HIGHEST CONC = 23 12/17 15:07  
LAST CALIBRATED :02/15 14:33

MODULE 2 %LEL COMB

LOW ALARM 12/10 12:08  
HIGH ALARM 00/00 00:00  
LOWEST CONC = 00 12/16 11:36  
HIGHEST CONC = 45 12/10 12:35  
LAST CALIBRATED :02/16 11:35

MODULE 3 %LEL COMB

LOW ALARM 12/10 11:57  
HIGH ALARM 12/10 12:20  
LOWEST CONC = 00 03/16 10:59