

<SYSTEM REPORT>

ID:A USER # = 1 12/17 15:08

M5000 VER:4.19

PN 8C

CALIBRATION CONFIGURATION:

LC: OFF CAL FREQ= 90 DAYS

H2S = 20 CAL DUE = 88 DAYS

COMB= 60 CAL DUE = 89 DAYS <SEL>

PRINTER CONFIGURATION:

PRINTER: ON STATUS = 07 DAYS HIST:YES

H2S : MIN. CONC. = 05 CHANGE = 03

COMB: MIN. CONC. = 08 CHANGE = 13

MODULE CONFIGURATION :

MODULE 1 IS ON H2S CHANNEL=1

LA= 10 NLATCH HA= 20 LATCH

FACTOR(S)= 100 150 200

MODULE 2 IS ON COMB CHANNEL=2

LA=2 0 NLATCH HA= 60 LATCH

FACTOR = 100