


**MANUFACTURERS DECLARATION OF CONFORMITY****Product identification:**

Model: 5100-28-IT
Description: Infra Red Gas Sensor Module – IT Series

Manufacturer:

Sierra Monitor Corporation
1991 Tarob Court
Milpitas, CA 95035 USA

The product has been tested by:	Sira Certification Service (Notified Body #0518) Rake Lane, Eccleston, Chester, CH4 9JN, England
Means of conformity to ATEX Pending  II2G Ex d IIB T6 Gb Ta= -20°C to +60°C (Directives and or standards used)	ATEX Directive 94/9/EC EN 60079-0:2012 EN 60079-1:2007

Means of conformity

We declare that the product listed above is in conformity and complies with the essential requirements based on test results using harmonized standards in accordance with directives mentioned.

Signature of representative / manufacturer:



Michael C. Farr
Vice President, Operations
Sierra Monitor Corporation
1991 Tarob Court Milpitas, CA 95035 USA